



SYNCHRONIZATION OF COUNTERSHOCK IN VENTRICULAR DEFIBRILLATION

CARLISLE, E.J.F., Kernohan, G., ALLEN, J.D., BAILEY, A., ANDERSON, J., & ADGEY, A.A.J. (1986). SYNCHRONIZATION OF COUNTERSHOCK IN VENTRICULAR DEFIBRILLATION. *IRISH JOURNAL OF MEDICAL SCIENCE*, 155(9), 327.

[Link to publication record in Ulster University Research Portal](#)

Published in:
IRISH JOURNAL OF MEDICAL SCIENCE

Publication Status:
Published (in print/issue): 01/09/1986

Document Version
Author Accepted version

General rights
Copyright for the publications made accessible via Ulster University's Research Portal is retained by the author(s) and / or other copyright owners and it is a condition of accessing these publications that users recognise and abide by the legal requirements associated with these rights.

Take down policy
The Research Portal is Ulster University's institutional repository that provides access to Ulster's research outputs. Every effort has been made to ensure that content in the Research Portal does not infringe any person's rights, or applicable UK laws. If you discover content in the Research Portal that you believe breaches copyright or violates any law, please contact pure-support@ulster.ac.uk.

SYNCHRONIZATION OF COUNTERSHOCK IN VENTRICULAR DEFIBRILLATION

E.J.F. Carlisle, W.G. Kernohan, J.D. Allen, A. Bailey, J. Anderson, and A.A.J. Adgey

Regional Medical Cardiology Centre, The Royal Victoria Hospital, Departments of Physiology and Orthopaedic Surgery, Queens University, and Department of Electrical Engineering, University of Ulster

Defibrillation of the heart occurs when a critical mass of myocardium is depolarised simultaneously, usually by electrical means. If the peaks and troughs of ventricular fibrillation (VF) indicate differing ratios of depolarisation to repolarisation in the heart, then the minimum energy required for defibrillation might be different for shocks synchronized to either the peaks or the troughs of the VF waveform. Lowering the energy required for defibrillation would minimise the chance of myocardial damage due to the countershocks themselves.

12 dogs were anaesthetised with pentobarbitone and ventilated. Arterial blood pressure, pH, and the ECG were monitored. VF was induced electrically via a pacing wire in the right ventricle. A custom built synchronization circuit sensed the peaks of the VF waveform. After a preset delay it triggered the defibrillator discharge so that the shock was delivered either at the following trough or the next peak.

Is VF a random arrhythmia? Spectral analysis of VF showed a periodic component in the arrhythmia, suggesting that synchronization of defibrillation might be possible. The first 40 seconds of VF (10 dogs) showed a dominant frequency of 11.48 ± 0.54 Hz (mean \pm SEM) with -3 db limits of 10.64 ± 0.54 Hz (lower) and 12.21 ± 0.51 Hz (upper).

The ventricular defibrillation threshold was determined as the smallest transthoracic shock to defibrillate the heart. Each experiment was divided into 3 runs and in each run, the ventricular defibrillation threshold was found either for transthoracic shocks delivered to the peaks of VF, to the troughs of VF, or for shocks delivered in unsynchronised mode.

There was no significant difference in the threshold delivered current or threshold delivered energy for shocks synchronized to the peaks of VF, to the troughs, or for unsynchronized shocks.

This study supports the previous reports of some periodicity in the VF waveform. However there appears to be no advantage in using synchronized transthoracic countershocks in VF.